

- | | |
|--|--|
| LD1 ~ Cup valve | LD23 ~ Decontaminant bottle empty switch |
| LD2 ~ Bowl valve | LD24 ~ Air for distil water bottle |
| LD3 ~ Large suction valve | LD25 ~ City water switch |
| LD4 ~ Small suction valve | LD26 ~ Durr valve input |
| LD5 ~ Bottle air valve | LD27 ~ Cup button |
| LD6 ~ City water valve | LD28 ~ Patient cup button |
| LD7 ~ Durr valve | LD29 ~ Bowl button |
| LD8 ~ Cattani rinse valve | LD30 ~ Bowl end switch Left |
| LD9 ~ Bowl - decontamination valve | LD31 ~ Bowl switch middle |
| LD10 ~ Syringe valve | LD32 ~ Bowl end switch right (Patient) |
| LD11 ~ Large suction - water for decontamination valve | LD33 ~ Water heating relay ON |
| LD12 ~ Small suction - water for decontamination valve | LD34 ~ Door open relay |
| LD13 ~ Large suction - Decontaminant for decontamination valve | LD35 ~ Suction relay ON |
| LD14 ~ Small suction - Decontaminant for decontamination valve | LD36 ~ Cattani empty |
| LD15 ~ Cup - disinfection valve | LD37 ~ +34V |
| LD16 ~ Tools water - disinfection valve | LD38 ~ Heartbeat - System running indication |
| LD17 ~ Large suction pick up switch | LD39 ~ +24V |
| LD18 ~ Small suction pick up switch | LD40 ~ CAN +5V |
| LD19 ~ Safety switch | LD41 ~ Curing Light ON |
| LD20 ~ Syringe pick up switch | LD42 ~ Ejector valve |
| LD21 ~ Xenos sensor ON | LD43 ~ Desinfectant bottle empty switch |
| LD22 ~ Error LED | LD44 ~ Distil water bottle empty switch |
| | LD45 ~ +3.3V |